

For Research Use Only

FlexAble FITC Plus Antibody Labeling Kit for Rat Kappa Light Chain

Catalog Number: KFA126



产品介绍

Product name	FlexAble FITC Plus Antibody Labeling Kit for Rat Kappa Light Chain
Assay type	Antibody labeling
Tested applications	FC
Species Reactivity	Rat Kappa Light Chain
Antibody amount per labeling reaction	0.5 µg antibody
Conjugate	FITC Plus
Excitation / Emission maxima wavelengths	495 nm / 524 nm

产品成分

Component	10 rxns	50 rxns	4×50 rxns
FITC Plus FlexLinker for Rat Kappa Light Chain	10 µL	50 µL	4×50 µL
FlexQuencher for Rat Kappa Light Chain	20 µL	100 µL	4×100 µL
FlexBuffer	100 µL	500 µL	4×500 µL

包装规格

10/50/4×50 reactions

保存条件

Store for 1 year at -20°C or for 6 months at +4°C after shipment. Avoid exposure to light.

其他

Q: What are the FlexLinker, FlexQuencher and FlexBuffer?

A: The FlexLinker is a small polypeptide to which dyes are covalently conjugated that can label unconjugated primary antibodies. The FlexQuencher is an Fc-containing fragment that neutralizes the excess FlexLinker. The FlexBuffer is a PBS-based buffer.

Q: What is the largest quantity I can label?

A: With a standard kit size (50 reactions), you can label 25 µg of one antibody or up to 50 different antibodies. You can easily scale up the antibody amount per labeling approach.

Q: What is the lowest concentration of my primary antibody that I can use?

A: Our protocol uses 0.5 µg of primary antibody in 7 µL, which ends up at 0.07 mg/mL. If the concentration of your antibody is lower, you can also use a larger volume than 7 µL.

Q: Can I label primary antibodies stored in BSA, glycerol, Tris buffer and/or preservatives?

A: Yes, FlexAble Antibody Labeling Kits have been validated with carriers and amine buffers. Neither BSA nor amine buffers, in any chosen concentration, interfere with the labeling. 50% glycerol as well as preservatives like sodium azide are also compatible with the kit.

Q: How many different primary antibodies can I label with one kit?


A: You can label up to 50 different antibodies with our FlexAble 50 rxn Kit, and up to 10 antibodies with our FlexAble 10 rxn Kit.

Q: Will I observe cross-reactivity/leaking when I use two FlexAble-labeled antibodies from the same species during multiplexing?


A: FlexAble labels primary antibodies with a high affinity FlexLinker. Dissociation of FlexLinker from one antibody and association to another antibody is rare. If you observe leaking, we recommend adding more FlexQuencher to remove unbound FlexLinker, or you can try sequential staining of the labeled antibodies.

[More FAQs](#)


Validation Data



Human peripheral blood lymphocytes (PBMC) were either un-stimulated (left) or stimulated with PMA and Ionomycin for 6 hours in the presence of monensin (right). Cells were fixed and permeabilized with Fcγ3/Transcription Factor Staining Buffer Kit (PF00011). After wash, cells were stained with rat anti-human TNF- α ... antibody labeled with FlexAble FITC Plus for Rat Kappa Light Chain (KFA126) and CD3 antibody (CL750-65151).



1X10⁶ C57BL/6 mouse splenocytes were surface stained with 0.5 μ g rat anti-Mouse CD4 (65104-1-Ig, Clone: GK1.5) labeled with FlexAble FITC Plus Kit (KFA126) and 0.5 μ g rat anti-Mouse CD3 (65077-1-Ig, Clone: 17A2) labeled with FlexAble CoraLite Plus 647 Kit (KFA123). Cells were not fixed.



1X10⁶ of LPS treated mouse splenocytes were surface stained with rat anti-Mouse CD86 (65068-1-Ig, Clone: GL1, red) or rat IgG2a isotype control (65209-1-Ig, blue) labeled with FlexAble FITC Plus for Rat Kappa Light Chain (KFA126). Cells were not fixed.

For technical support and original validation data for this product please contact

T: 027-87531629

E: Proteintech-CN@ptglab.com

W: ptgcn.com

This product is exclusively available under Proteintech Group brand and is not available to purchase from any other manufacturer.